

Declaration of conformity according to Regulation (EU) 2017/745

We,

**ARNOWA GmbH
Kugelbreite 30
33154 Salzkotten
Germany**

declare under our sole responsibility that the products listed below meet all provisions of Regulation (EU) 2017/745 on medical devices.

The products are class I medical products in accordance with Annex VIII, Chapter III, and meet the applicable "General Safety and Performance Requirements" according to Annex I of Regulation (EU) 2017/745.

Conformity assessment procedure (EU) 2017/745: Annexes II and III

Basic UDI-DI: 426059752ARNOMEDPADS5U

Applied Standards: ISO 10993-5:2009, ISO 10993-10:2013

Intended purpose: A flat piece of absorbent fabric material [e.g. nonwoven (PP), polyethylene (PE), cellulose] intended to be placed on a surface (e.g. bed mattress, table, floor) and under a patient to absorb body fluids generated during surgical procedures, childbirth and/or incontinence; it does not contain antimicrobial agents. It is intended for use in the healthcare facility and at home. It is a non-sterile disposable product.

Validity: If no changes are made to the product, the declaration is valid for 5 years.

Product name	Art.-No.	Sizes	Colour
ARNOMED FLOCKPAD SUPERDRY	623-60x40	60 x 40 cm	Blue
ARNOMED FLOCKPAD SUPERDRY	623-60x60	60 x 60 cm	Blue
ARNOMED FLOCKPAD SUPERDRY	623-60x90	60 x 90 cm	Blue
ARNOMED FLOCKPAD SUPERDRY LIGHT	624-60x90	60 x 90 cm	Blue

Salzkotten, 04.06.2024



Daniel Müller
Managing Director
ARNOWA GmbH

File: Declaration of Conformity ARNOMED FLOCKPAD	Revision/Version: V1	Product: 426059752ARNOMEDPAD S5U
Copyright: ARNOWA	Applicable: TD-ENG_Declaration of Conformity ARNOMED FLOCKPAD	Page: - 1 -