



easyCARE nitrile

Nitrile examination gloves, powder-free, non-sterile

TD-49-I.2.c-2.1.1
revision 9
date: 2021-05-10
amendment of: 2025-11-20

MANUFACTURER: ZARYS International Group sp. z o.o. sp.k.
ADDRESS: ul. Pod Borem 18, 41-808 Zabrze, Poland; T: +48 32 376 07 00-02,06 F: +48 32 370 38 08; www.zarys.com

CLASSIFICATION Class I medical device compliant with Medical Devices Regulation (EU) 2017/745 (MDR)
 Personal Protective Equipment Category III according to PPE Regulation (EU) 2016/425
 The information provided by the manufacturer with the product meets the requirements of Annex I to Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 and of point 1.4 of Annex II to Regulation (EU) 2016/425 of the European Parliament and of the Council of 9 March 2016

INTENDED PURPOSE

- the device is intended for patient diagnostics, providing a barrier to the transmission of microorganisms between the operator and the patient, minimizing the risk of cross-contamination

FEATURES

- raw material:** 100% nitrile (acrylonitrile butadiene)
- colour:** blue
- cuff:** beaded
- powder:** powder-free, the content of powder residue below 2mg per glove
- AQL:** 1,0
- internal surface:** chlorinated and polymerized
- outer surface:** polymerized, micro-textured throughout the glove with additional texture at the fingertips
- compliance with the standards:** EN 455-1:2020+A2:2024, EN 455-2:2024, EN 455-3:2023, EN 455-4:2009, EN ISO 13485:2016, EN ISO 13485:2016/A11:2021, EN ISO 9001:2015, EN ISO 9001:2015/A1:2024, EN ISO 14001:2015, EN ISO 14001:2015/A1:2024, EN ISO 14971:2019, EN ISO 14971:2019/A11:2021, EN 62366-1:2015, EN 62366-1:2015/A1:2020, EN ISO 15223-1:2021, EN ISO 20417:2021, EN ISO 10993-1:2020, EN ISO 10993-5:2009/A11:2025, EN ISO 10993-10:2023, EN ISO 10993-23:2021, EN ISO 11737-1:2018, EN ISO 11737-1:2018/A1:2021, ASTM F1671, ASTM D6978, ASTM F739, EN ISO 21420:2020, EN ISO 21420:2020/A1:2024, EN ISO 374-1:2016, EN ISO 374-1:2016/A1:2018, EN ISO 374-2:2019, EN 16523-1:2015+A1:2018, EN ISO 374-4:2019, EN ISO 374-5:2016, ASTM D6319
- allowed for contact with food:** in accordance with Regulation (EC) 1935/2004 and Commission Regulation (EC) 10/2011
- intermediate packaging:** 100 pcs or 200 pcs with a colour-coded packaging for easy size recognition
- size:** XS-XL
- single-use**
- non-sterile**
- shelf life:** 5 years



CATALOGUE NUMBER	SIZE	INTERMEDIATE PACKAGING	BULK PACKAGING
RNBXS10001	XS	100 pcs	10 x 100 pcs
RNBS10001	S	100 pcs	10 x 100 pcs
RNBM10001	M	100 pcs	10 x 100 pcs
RNBL10001	L	100 pcs	10 x 100 pcs
RNBXL10001	XL	100 pcs	10 x 100 pcs
CATALOGUE NUMBER	SIZE	INTERMEDIATE PACKAGING	BULK PACKAGING
RNBXS20001	XS	200 pcs	10 x 200 pcs
RNBS20001	S	200 pcs	10 x 200 pcs



easyCARE nitrile
Nitrile examination gloves, powder-free,
non-sterile

TD-49-I.2.c-2.1.1

revision 9

date: 2021-05-10

amendment of: 2025-11-20

RNBM20001	M	200 pcs	10 x 200 pcs
RNBL20001	L	200 pcs	10 x 200 pcs
RNBXL20001	XL	200 pcs	10 x 200 pcs